


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVRL-1-PS*  
S.N.: *9265*  
Date: *MAR 15 2000*



- a) Output signal amplitude:  
*-200 V (70 50Ω)*
- b) Pulse width:  
*5 TO 100 ns*
- c) Rise time:  
*≤ 2 ns*
- d) Fall time:  
*≤ 3 ns*
- e) PRF:  
*1 TO 5 KHz*
- f) Jitter, stability:  
*60%*
- g) Prime power:  
*100 ± 10%*

*60 db ATTN*  
*100 V OUTS / DIV*  
*PRF = 3 KHz*

*50-60 Hz*  
