


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-1M2-P

S.N.: 9270

Date: Feb 25 2000



- a) Output signal amplitude:  $\approx +60 \text{ Volts (50\%)} -$
- b) Pulse width:  $\leq 2 \text{ ns} -$
- c) Rise time: NA -
- d) Fall time: NA -
- e) PRF: 0 to 1.0 MHz -
- f) Jitter, stability: OK -
- g) Prime power:  $+135 \pm 5 \text{ VDC}$   
 $3 \text{ mA MAX.}$

50  $\Omega$  ATTEN  
-  $\approx 32 \text{ V/DIV}$   
PRF  $\approx 300 \text{ kHz}$

*[Signature]*