


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-1M2-P*
S.N.: *9271*
Date: *FEB 25 2000*



- a) Output signal amplitude: ✓
≥ +60 Volts (50Ω)
- b) Pulse width:
≤ 2 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:
+135 ± 5 VDC
3 mA MAX.

50 Ω 1ATTN
32 V/DIV
PRF = 300 kHz

ef