


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-1M2-P

S.N.: 9272

Date: Feb 25 2000



a) Output signal amplitude:  $\geq +60 \text{ Volts (50\%)} \checkmark$

b) Pulse width:  $\leq 2 \text{ ns} \checkmark$

c) Rise time: NA  $\checkmark$

d) Fall time: NA  $\checkmark$

e) PRF: 0 to 1.0 MHz  $\checkmark$

f) Jitter, stability: OK  $\checkmark$

g) Prime power:  $+135 \pm 5 \text{ VDC}$   
3 mA max.

50  $\Omega$  ATTEN  
12 32 V/DIV  
PRF = 300 kHz.