


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-m2-P*

S.N.: *9274*

Date: *FEB 25 2000*



- a) Output signal amplitude: ✓
≥ +60 Volts (50Ω)
- b) Pulse width: ✓
≤ 2 ns
- c) Rise time: ✓
NA
- d) Fall time: ✓
NA
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: ✓
OK
- g) Prime power: ✓
+135 ± 5 VDC
3 mA MAX.

50 Ω 6 ATTEN
≈ 32 V/DIV
PRF ≈ 300 kHz

[Signature]