


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-14B-C-P-PN-05-EA*

S.N.: *9275*

Date: *APRIL 24 2000*



- a) Output signal amplitude:
0 TO ± 40 VOLT (TO 50V)
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 NS
- d) Fall time:
≤ 250 NS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 MHz

50 dB ATTEN
∴ ≈ 32 VOLT/DIV
PRF ≈ 100 KHz