


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-MPT-EGGID

S.N.: 909A

Date: FEB 7 2000



- a) Output signal amplitude:  $\approx +120$  VOLTS (70.50V)
- b) Pulse width: 600 ps (FWHM)
- c) Rise time:  $\leq 150$  ps
- d) Fall time:  $\leq 275$  ps
- e) PRF: 0 TO 2.0 MHz
- f) Jitter, stability: OK
- g) Prime power:

40 db ATTN

10 VOLTS/DIV

TOP: 500 ps/DIV

BOT: 100 ps/DIV (RISE TIME)

+15 VDC

250 mA  
MAX

PRF  $\approx$  1.0 MHz.