


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-M-1-CT-M*

S.N.: *9295*

Date: *MAY 5 2000*



- a) Output signal amplitude:
0 to 70V
- b) Pulse width:
0.2 to 4.0 ns
- c) Rise time:
 $\le 100ps$
- d) Fall time:
 $\le 135ps$
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]