


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3A-C-F*
S.N.: *9298*
Date: *MAR 28 2000*



- a) Output signal amplitude:
0 to +450 (70.5V)
- b) Pulse width:
≤ 2 ns (20% rise point)
- c) Rise time:
~ 0.8 ns
- d) Fall time:
≤ 0.8 ns
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
125/240V

60 db return
100 VOLT/DIV
PRF = 20 kHz

50-60 Hz
[Signature]