


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AU-HV3-P-C*

S.N.: *9299*

Date: *MAR 28 2000*



- a) Output signal amplitude:
0 to +40 V (70.50 n)
- b) Pulse width:
0.4 to 2.0 ns
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/280 V
50-60 kHz

50 dB ATTEN
≈ 32 V/DIV

PRF = 100 kHz

[Signature]