


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVIR-1-C-N*

S.N.: *9300*

Date: *JUNE 8 2000*



*60 dB ATTN*

*100 V/DIV*

*PRF = 20 KHz*

- a) Output signal amplitude:  
*0 TO -200 VOLTS*
- b) Pulse width:  
*1 TO 5 NS*
- c) Rise time:  
*≤ 0.5 NS*
- d) Fall time:  
*0.5 - 1.0 NS*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50 - 60 Hz*