


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-MP-1*

S.N.: *9311*

Date: *MAR 10 2000*



- a) Output signal amplitude:  
*> +40 (TO 50Ω)*
- b) Pulse width:  
*2 NS TO 100 NS*
- c) Rise time:  
*≤ 350 ps*
- d) Fall time:  
*≤ 2 NS*
- e) PRF: *PW = 2 NS: 0 TO 1.0 MHz*  
*PW = 100 NS: 0 TO 20 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *+15 VDC*  
*150 mA*  
*(20 A)*

*50 Ω ATTN.: = 32 VOLTS/DIV*

*TOP: DL = 24"*

*BOT: DL = 0"*

*[Signature]*