


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1423-C-PW*

S.N.: *9325*

Date: *JUNE 1 2000*



- a) Output signal amplitude:
0 to ± 40 VOLTS
- b) Pulse width: *(to 50%)*
0.4 to 2.0 NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

50 dB ATTEN

232 VOLTS/DIV

PRF = 100 KHz