


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANM-2-C-PN-D-M*

S.N.: *9332*

Date: *JUNE 23 2000*



- a) Output signal amplitude:  
*0 TO +15V (TO 50V)*
- b) Pulse width:  
*0.2 TO 2.0NS*
- c) Rise time:  
*≤ 100ps*
- d) Fall time:  
*≤ 135ps*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*

*40 db ATTN*  
*10 V/DIV*  
*PRF ≈ 12 MHz*

*50 60 Hz*  
*[Signature]*