


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVMP-3-P*  
S.N.: *9344*  
Date: *MAY 26 2000*



- a) Output signal amplitude:  
*0 TO +20 V*
- b) Pulse width:  
*8 TO 100 NS*
- c) Rise time:  
*≤ 200 ps*
- d) Fall time:  
*≤ 300 ps*
- e) PRF:  
*0 TO 100 kHz*
- f) Jitter, stability:

g) Prime power: *OK*  
*+24 VDC*  
*250 mA (MAX)*

*40dB ATTN*  
*10 V/DIV*  
*TDR 5 NS/DIV*  
*BOT 500ps/DIV (RISE TIME)*  
*PRF = 100 kHz.*