


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMP-3-A*  
S.N.: *9347*  
Date: *MAY 26 2000*



- a) Output signal amplitude:  
*0 to +20 VOLT (to 50V)*
- b) Pulse width:  
*8 to 100 ns*
- c) Rise time:  
*≤ 200 ps.*
- d) Fall time:  
*≤ 300 ps.*
- e) PRF: *0 to 1 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 VDC,*  
*250 mA*  
*(MAX)*

*40 dB ATTEN*  
*10 VOLT/DIV*  
*TOP 5 NS/DIV*  
*BOT 500 ps/DIV (RISE TIME)*

*PRF = 100 kHz*

*[Signature]*