


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMR-2-P*

S.N.: *9343*

Date: *MAY 26 2000*



- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width: *(to 50%)*
20 TO 200N
(20% MAX DUTY CYCLE)
- c) Rise time:
≤ 3NS
- d) Fall time:
≤ 3NS
- e) PRF: *0 TO 10 MHz*
(20% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC
500 mA MAX.