


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-MV-HUB-C-P-ET-EN*

S.N.: *9360*

Date: *JULY 18 2000*



- a) Output signal amplitude:
0 TO +40 V
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*

50 db ATTN
≈ 32 VOLTS / DIV
PRF = 100K Hz

- g) Prime power:
120/240 V
50-60 Hz

[Signature]