


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AU-14V3-C-PN

S.N.: 9372

Date: JUNE 14 2000



- a) Output signal amplitude:
0 TO ± 40 VOLTS
- b) Pulse width: ($R_L = 50 \Omega$)
0.4 TO 4.0 NS
- c) Rise time:
 ≤ 150 ps
- d) Fall time:
250 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

50 db ATTN
 ≈ 32 VOLTS/DIV

PRF ≈ 100 KHz

[Signature]