


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-3-PS-LTMA-D-T*

S.N.: *9420*

Date: *AUG 17 2000*



- a) Output signal amplitude: *0 TO +40 VOLTS*
- b) Pulse width: *(TO 50 n)*
500ps
- c) Rise time: *≤ 200ps*
- d) Fall time: *≤ 200ps*
- e) PRF: *0 TO 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

TIP INPUT TRIGGER
20 mV/DIV
1 NS/DIV

BOT OUTPUT
20 VOLTS/DIV
1 NS/DIV
10 MHz PRF