


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-1-C-OT-M-P*

S.N.: *9425*

Date: *JULY 13 2000*



- a) Output signal amplitude: *0 TO 75 VOLTS*
- b) Pulse width: *(TO 50ns)*
0.2 TO 1.0 NS
- ① c) Rise time: *≤ 100ps*
- d) Fall time: *≤ 100ps*
- ② e) PRF: *50 TO 250 MHz*
- ③ f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

40 db ATTEN.: 5 V/DIV

① *PRF ≈ 100 MHz.*
2 NS / DIV

② *PRF ≈ 100 MHz*
500 ps / DIV

③ *PRF ≈ 200 MHz.*
1 NS / DIV.

[Signature]