


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-P*  
S.N.: *9427*  
Date: *JULY 18 2000*



- a) Output signal amplitude:  
*0.70 to 1.5 VOLTS*
- b) Pulse width:  
*0.2 to 1.0 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 100 PS*
- e) PRF:  
*50 TO 250 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*+24 VDC,  
400 mA  
MAX.*

*40 dB ATTEN  
5 VOLTS DIV  
≈ 125 MHz.*