


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2-C-BE03-R5-P-0P1-M

S.N.: 9428

Date: JUNE 29 2000



- a) Output signal amplitude:
0 TO +2 AMP ✓
- b) Pulse width: ✓
17 NS
- c) Rise time: ✓
≤ 2 NS
- d) Fall time: ✓
≤ 2 NS
- e) PRF: ✓
0 TO 20 kHz
- f) Jitter, stability: ✓
OK
- g) Prime power:
120/240V, 50-60 Hz

MAINFRAME OUT TO 60 dB
100V/DIV, 2 AMP/DIV
5 NS/DIV

[Signature]

PRF = 20 KHz.
IN SERIES WITH LOAD.
REF = 20 KHz