

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-C-PN*

S.N.: *9448*

Date: *Jul 22 2000*


- a) Output signal amplitude:
0 TO ± 200 V
- b) Pulse width: *($R_L \geq 50\Omega$)*
0.1 TO 100 μ s
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability: *(0.5% MAX DUTY CYCLE)*
OK
- g) Prime power: *120/240 V*
50-60 Hz



9448


NARROW PULSE

$R_L = 50 \Omega$



9448

WIDE PULSE
 $R_L = 50\Omega$



B-7 no.

22 Aug 2000
14:59:43