

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-3-CPN

S.N.: 9449

Date: AUG 22 2000


- a) Output signal amplitude:
0 to ± 200 V ($R_L \geq 50 \Omega$)
- b) Pulse width:
0.1 to 100 μ s
(0.5% max duty cycle)
- c) Rise time:
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF: 0 to 10 kHz
0.5% max duty cycle
- f) Jitter, stability:
OK
- g) Prime power: 120/240 V
50-60 Hz.



9449

NARROW PULSE

$R_L = 50\Omega$



9449

WIDE PULSE

$R_L = 50\Omega$

