


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-PS-P*

S.N.: *9450*

Date: *JULY 11 2000*



- a) Output signal amplitude:
0 TO +5V (7050Ω)
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF: *0 TO 25 MHz.*
- f) Jitter, stability: *OK.*
- g) Prime power: *120/240V*

40 dB ATTN
-5 VOLTS / DIV
PRF ≈ 10 MHz.

50-60 MHz.