


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-1-C-P*

S.N.: *9453*

Date: *SEPT 22 2000*



- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *UPD 50ns*
 - a) *0.2 TO 5.0ns*
 - b) *5.0 TO 100ns*
- c) Rise time: *≤ 100 ps*
- d) Fall time: *≤ 135 ps*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50 160 Hz