


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-3A-P5-F*
S.N.: *9474*
Date: *AUG 31 2000*



- a) Output signal amplitude: *0 TO +300V (TO 500V)*
- b) Pulse width: *5 TO 100ns*
- c) Rise time: *≤ 600ps*
- d) Fall time: *≤ 5ns*
- e) PRF: *0 TO 5KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

60 dB ATTEN
- 100 V/DIV
TDR 10 ns/DIV
BST 1 ns/DIV
(RISE TIME)
PRF = 2 KHz