


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3A-C-T
S.N.: 9475
Date: AUG 31 2000



- a) Output signal amplitude: 0 TO +300 V (TO 50%)
- b) Pulse width: 5 TO 100 NS
- c) Rise time: ≤ 600 ps
- d) Fall time: ≤ 5 ns
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 Hz

60 dB ATTEN
100 VOLTS/DIV
TOP 10 NS/DIV
BOT 1.0 NS/DIV
(RISE TIME)

PRF = 2 KHz

[Signature]