


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-MN-1-C-OT-P*

S.N.: *9432*

Date: *AUG 3 2000*



- a) Output signal amplitude:
0 TO +10 V (to 50Ω)
- b) Pulse width:
0.2 TO 4.0 NS.
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/240 V*
50 to 1 Hz

40 dB ATTEN

TOP 1 NS / DIV
+5 Vdc OFFSET
PW NORM 1 MHz

BOT 1 NS / DIV
NO OFFSET
PW AT 1 MHz
PRF ≈ 100 kHz