


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANH-MPT-EGGIE

S.N.: 9436

Date: Aug 15 2000



a) Output signal amplitude:
10 V (70 50%)

b) Pulse width:
1.5 ns FWHM

c) Rise time:
≤ 100 ps

d) Fall time:
≈ 3 ns

e) PRF: 0 to 2 MHz

f) Jitter, stability:
OK

g) Prime power:
a) +15 VDC
160 mA (max)

b) -5 VDC
10 mA

40 db ATTN
10 V / DIV
TOP 1 ns / DIV
BOT 100 ps / DIV (RISE TIME)

PRF = 2.0 MHz

(LARGE JITTER DUE TO TRIGGER SOURCE)