


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ALH-MPT-EGGIE*

S.N.: *9502*

Date: *NOV 9 - 2008*



- a) Output signal amplitude:
≥ +10 VDC (70583)
- b) Pulse width:
1.5 NS FWHM
- c) Rise time:
≤ 100 ps
- d) Fall time:
≈ 3 NS
- e) PRF:
0 TO 2 MHz
- f) Jitter, stability:
OK
- g) Prime power:
 - a) *+15 VDC*
160 mA
 - b) *-5 VDC*
10 mA

460 db ATTEN
10 V/DIV
TOP: 1 NS/DIV
BOT: 100 ps/DIV

PRF = 2.0 MHz