


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-EBBIE*

S.N.: *9514*

Date: *DEC 21 - 2008*



- a) Output signal amplitude: $\geq +10 \text{ V}_{\text{pk}}$ (705.5)
- b) Pulse width: 1.5 ns FWHM
- c) Rise time: $\leq 100 \text{ ps}$
- d) Fall time: $\approx 3 \text{ ns}$
- e) PRF: $0 \text{ TO } 2 \text{ MHz}$
- f) Jitter, stability: OK
- g) Prime power:
 - a) $+15 \text{ VDC}$ - 160 mA
 - b) -5 VDC - 10 mA

40 db ATTEN
10 V/DIV
TOP: 1 ns/DIV
BOT: 100 ps/DIV
PRF = 2.0 MHz

[Signature]