


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-4A-C-N-M*

S.N.: *9530*

Date: *OCT 24 2000*



- a) Output signal amplitude:
0 TO -650 VOLTS
- b) Pulse width: *(70 50%)*
≤ 5 NS (20% RISE TIME)
- c) Rise time:
≤ 1.5 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 160 Hz

60 db ATTEN
100 V/DIV
5 NS/DIV

PRF = 10.0 KHz.