


PULSE GENERATOR
PERFORMANCE CHECK

Model: *BVM-1-C-A*
S.N.: *9538*
Date: *NOV 7 2000*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width: *(TO 50N)*
0.2 TO 6.0 NS
- c) Rise time:
 $\le 100 ps$
- d) Fall time:
 $\le 135 ps$
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN
5 V/DIV
PRF $\approx 10 MHz$

120/240V
50-60 Hz
[Signature]