


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANP-AN-423-PS-F

S.N.: 9557

Date: OCT 19 2000



- a) Output signal amplitude:
0 TO +40V (TO 50V)
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
 $\leq 150 \text{ ps}$
- d) Fall time:
 $\leq 250 \text{ ps}$
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50/60Hz

50 dB ATTN
32 VOLTS/DIV
1 NS/DIV
300 KHz