


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PM-2-P-D-M*

S.N.: *9558*

Date: *OCT 18 2000*



- a) Output signal amplitude:  
*0 TO +15V (TO 50n)*
- b) Pulse width:  
*0.2 TO 2-0NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 VDC,*  
*500 mA (MAX)*

*40 db ATTN*

*-10 VOLTS/DIV*

*PRF = 10 MHz*

*[Signature]*