


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MH-HVI-EA5-M1*

S.N.: *9573*

Date: *DEC 12 2000*



- a) Output signal amplitude: *0 TO +100 V*
- b) Pulse width: *CTD 50ns*
≤ 1 NS
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 200 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*
250 mA (max)

60 db ATTN

100 V 0.83 / DIV
PRF = 200 KHz