


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMM-2-N GA-EW

S.N.: 9585

Date: JAN 25 2001



- a) Output signal amplitude:
0 TO -5V (0.50V)
- b) Pulse width:
1 TO 10 NS
- c) Rise time:
 ≤ 300 ps
- d) Fall time:
 ≤ 600 ps
- e) PRF: 0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
5 VOLTS/DIV.
PRF = 10 MHz

+24 VDC
500 mA (MAX)