


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1-P*
S.N.: *9589*
Date: *DEC 4 2000*



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
(TO 50N)
0.2 TO 4.0N
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 35 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC
250 mA (MAX)

40 db ATTN
.10 VOLTS/DIV
PCRF = 100 KHz
TOP 1.0 ns/DIV
BOI 100 ps/DIV
(RISE TIME)