


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3C-C-P
 S.N.: 9594
 Date: JAN 12 2001



- a) Output signal amplitude: 0 to +120V
- b) Pulse width: (to 50%)
3 to 100ns
- c) Rise time: ≤ 300 ps
- d) Fall time: ≤ 3.0 ns
- e) PRF: 0 to 5 kHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V

60 db ATTEN
100 VOLT/DIV
10 ns/DIV
BAT 1 ns/DIV (RISE TIME)
PRF = 5 kHz
(Rc = 50n)

50-60 MHz
