


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-BB-C-PN*

S.N.: *9600*

Date: *DEC 9 2000*



- a) Output signal amplitude:
0 to +450 Volts
- b) Pulse width: *(to 50%)*
≤ 2 ns
- c) Rise time:
≤ 0.8 ns
- d) Fall time:
≤ 0.8 ns
- e) PRF: *0 to 20 kHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60MHz

60 db ATTEN
100 VOLTS/DIV
PRF = 10 kHz

[Signature]