


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMM-2-PS-TR-N*

S.N.: *9610*

Date: *JAN 8 2001*



- a) Output signal amplitude:  
*0 TO -5 VOLTS (TO 50%)*
- b) Pulse width:  
*1 TO 10 NS*
- c) Rise time:  
*300ps TO 2 NS*
- d) Fall time:  
*600ps TO 2 NS*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 dB ATTEN - 5 VOLTS/DIV*

*100 ± 10 VOLTS*

*TR SWITCH IN ALL  
S POSITIONS.*

*SD 160 Hz*

*PRF = 10 MHz*