


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVM4-3-PS-LTMA-D-T


S.N.: 9628

Date: JAN 30 2001



- a) Output signal amplitude:
0 TO +40 V (70 50 Ω)
- b) Pulse width:
500 ps
- c) Rise time:
 ≤ 200 ps
- d) Fall time:
 ≤ 20 ps
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

① INPUT TR 16 GHz 20mV/DIV
1 ns/DIV



③ OUTPUT 20V/DIV
1 ns/DIV
10 MHz BW