


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-PS-PN-05*

S.N.: *9643*

Date: *FEB 1 2001*



- a) Output signal amplitude:
0 to ±50 Volts
- b) Pulse width: *(7050ns)*
2 to 100ns
- c) Rise time:
≤ 350ns
- d) Fall time:
≤ 2ns
- e) PRF: *0 to 20 kHz*
- f) Jitter, stability:
- g) Prime power: *OK*

120/240V
50-60Hz

[Signature]

50 db ATTEN
- 32 u/DIU
TOP 10 ns/DIU
150T 1 ns/DIU
(RISE TIME)
PRF ≈ 10 kHz