

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV-110B-PS*

S.N.: *9670*

Date: *MAR 1 2001*

- a) Output signal amplitude:  
 *$\pm 200V$  ( $R_L > 50K$ )*
- b) *GAIN:  $\times 1$  to  $\times 100$*
- c) Rise time:  
 *$\leq 1\mu s$  ( $V_{OP}$ )*
- d) Fall time:  
 *$\leq 1\mu s$  ( $V_{OP}$ )*
- e) PRF: *DC TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:  
*120/240V*  
*\*50-60 Hz*

9670

$R_L = 50K.$

$V_{LOAD}$

$V_{IN}$

