


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-C-PN*

S.N.: *9639*

Date: *MAR 20 2001*



- a) Output signal amplitude:
0 TO ± 50 VOLTS (TO 50Ω)
- b) Pulse width:
2 TO 100NS
- c) Rise time:
≤ 350 ps.
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 20 KHz.*
- f) Jitter, stability:
OK
- g) Prime power:

50 db ATTN

32 VOLTS/DIV

TDR 10 NS/DIV

*Prot. 500 ps/DIV.
(RISE TIME)*

120/240V

50-60 Hz

[Signature]