


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14-HA-P-T1*

S.N.: *9690*

Date: *MAR 14 2001*



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 0.6 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
100 VOLTS/DIV
PRF ≈ 30 KHz

+15 VDC
200 mA (max)
[Signature]