


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-HB-C-P*

S.N.: *9698*

Date: *FEB 9 2001*



- a) Output signal amplitude:
0 to +40V (crossed)
- b) Pulse width:
0.4 to 2.0ns
- c) Rise time:
≤ 150ps
- d) Fall time:
≤ 250ps
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
170/240V
50-60Hz

[Signature]