

PULSE GENERATOR
PERFORMANCE CHECK

Model: *MV-110B-PS*

S.N.: *9711*

Date: *MAR 1 2001*

- a) Output signal amplitude:
 $\pm 200V (R_L > 50K)$
- b) *GAIN: $\times 1$ TO $\times 100$*
- c) Rise time:
 $\leq 1\mu S (V_{op})$
- d) Fall time:
 $\leq 1\mu S (V_{op})$
- e) PRF:
DC TO 100KTR
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

9711
 $R_c = 50\text{ K}$

V_{LOAD}

V_{IN}

